

Part-I EVALUATION REPORT FOR Indent NO. 23292

Three bidders were participated in the Bid:

Sr. No.	Particulars	M/S. Machines Tools centre Kolkata	M/S. R and D Instrument services Chennai	M/S. Electronet Equipment Pvt. Ltd. Pune
1.	MSE Purchase Preference/Reservation based on Udyam Registration Certificate	NO (Not applied)	YES as Bidder is Micro Manufacturing MSME	YES as Bidder is Micro Manufacturing MSME
2.	Compliance to Technical Specification *(Detail report is given below)	Yes	Yes	No
3.	Commercial Status	Accepted	Accepted	Accepted
4.	Status	Technically and Commercially Accepted	Technically and Commercially Accepted	Commercially Accepted but Technically Rejected Bidder has not provided compliance to the specifications

***Technical specification compliance details**

Technical Specification compliance for Item no-1:-

Sr. No.	Parameter	Description	Compliance by bidders		
			M/S. Machines Tools centre Kolkata	M/S. R and D Instrument services Chennai	M/S. Electronet Equipment Pvt. Ltd. Pune
1.	Range	0-300 kg/cm ²	Yes	Yes	Yes
2.	Dial Size	88mm or greater than 88 mm	Yes	Yes	Yes
3.	Display	Graphical LCD Display or Large LCD Display(15mm height or greater than 15 mm height)	Yes	Yes	Yes
4.	Wetted Parts:	SS304	Yes	Yes	No
5.	Case Material	Aluminium	Yes	Yes	No
6.	Accuracy	0.05% FS or better	Yes	Yes	Yes
7.	Connection	1/4" bsp Port Bottom entry with BSP to NPT conversable fitting or ½ " NPT Bottom entry	Yes	Yes	Yes
8.	Display Units	At least kg/cm2, bar, psi, KPa	Yes	Yes	Yes
9.	Over Pressure Limit	1.3 times full scale pressure or better	Yes	Yes	Yes
10.	Calibration Cert. from NABL Accredited Lab	Yes	Yes	Yes	Yes

Latit Patil, SO/E, KAPP-3&4	H.K. Gajjar, SO/F, KAPP-3&4	SME(F), KAPP-3&4 25/12/2024
--------------------------------	--------------------------------	--------------------------------

Signature Not Verified

Digitally signed by BARUN KUMAR PATRA

Date: 2024.09.09 10:51:32 IST

Location: eProcurement System for Central PSUs